


America

CERTIFICATE

No. U8 126702 0001 Rev. 00

Holder of Certificate: **technotrans solutions GmbH**
 Scherl 10
 58540 Meinerzhagen
 GERMANY

Certification Mark:



C US

Product: **Control units
 (temperature control device)**


This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713338924-00

Date, 2024-12-17

Siemon

(Thorsten Siemon)



America

CERTIFICATE

No. U8 126702 0001 Rev. 00

Model(s):
teco ci 140 base 60 UL
teco ci 140 high 60 UL
teco ci 140 eco 60 UL
teco ci 160 base 60 UL
teco ci 160 high 60 UL
teco ci 160 eco 60 UL
teco ci 180 base 60 UL
teco ci 180 high 60 UL
teco ci 180 eco 60 UL

Brand Name: technotrans

Tested according to:
UL 61010-1:2012/R:2023-06
UL 61010-2-010:2019/R:2023-09
CSA C22.2 No. 61010-1:2012/U3:2023-06
CSA C22.2 No. 61010-2-010:2019

Parameters:
Operating voltage: 3~, 440-480V
Frequency: 50/60Hz
Power: 10.9kW